

Declaration of conformity

Batch

Declaration of conformity given by the supplier of product according to the standard EN 45014




Establishment TECHNOSKLO Ltd. Držkov 135 applied and using system of quality according to the DIN EN ISO 9001: 2000, certificate made out by TÜV Rhainland FRG

Sort of product:	1605/AS GRADUATED PIPETTE
Nominal volume:	5 ml
Division:	0,05 ml
Permitted volume error:	± 0,03 ml
Average value of volume:	4,997 ml
Standard deviation of measurement:	± 0,004 ml
Number of charge:	09 . 01
Testing instrument:	balance - production number: 613343; id. number: 34/00, thermometer - production number: 2259-2071-05; id. number: 35/05. Used standards are in accordance to standards of Czech Metrology Institute.

We are declaring to our exclusive responsibility, that product, which relates to this declaration responds to the parameters of standard: ČSN EN ISO 835:2007.

Date of edition: 20. 08. 2009

Confirmed by:


Manager of Quality

T 3989 / 8