

Declaration of conformity

Individual

Declaration of conformity given by the supplier of product according to the standard EN 45014



Establishment TECHNOSKLO Ltd. Držkov 135 applied and using system of quality according to the DIN EN ISO 9001: 2000, certificate made out by TÜV Rhainland FRG


Sort of product:	1567 / ASST	BURETTE
Nominal volume:	50	ml
Division:	0,1	ml
Permitted volume error:	± 0,05	ml
Average value of volume:	50,011	ml
Uncertainty of measurement:	± 0,009	ml
Number of charge:	0468	

Testing instrument: balance - production number: 613343; id. number: 34/00,
thermometer - production number: 2259-2071-05; id. number: 35/05.
Used standards are in accordance to standards of Czech Metrology Institute.

We are declaring to our exclusive responsibility, that product, which relates to this declaration responds to the parameters of standard: EN ISO 385:2005.

Date of edition: 20. 08. 2009

Confirmed by:


Manager of Quality