

# Checking record



Establishment TECHNOSKLO Ltd. Držkov 135 applied and using system of quality according to the DIN EN ISO 9 001: 2000, certificate made out by TÜV Rhainland FRG


Sort of product: **1595 / AS ONE MARK PIPETTE**  
Nominal volume: **10 ml**  
Permitted volume error: **± 0,02 ml**

Testing instrument: balance - production number: 613343; id number: 34/00,  
thermometer - production number: 2259-2071-05; id. number: 35/00.  
Used testing instruments was took up to standards instruments of Czech metrology institute.

We are declaring to our exclusive responsibility, that product, which relates to this declaration responds to the parameters of standard: ČSN EN ISO 648.

Date of edition: 20. 08. 2009

Confirm:

  
manager of quality