

Declaration of conformity

Batch

Declaration of conformity given by the supplier of product according to the standard EN ISO IC 17050-1



Establishment TECHNOSKLO s.r.o. Držkov 135 applied and using system of quality according to the DIN EN ISO 9001: 2015, certificate made out by TÜV Rheinland FRG

Sort of product: **632 435 170 725 BURETTE WITH STRAIGHT STOPCOCK, GLASS KEY, WITH SCHELLBACH STRIPE, CLASS AS**

Nominal volume:	50	ml
Division:	0,1	ml
Permitted volume error:	± 0,050	ml
Average value of volume:	50,012	ml
Standard deviation of measurement:	± 0,006	ml

Testing instrument: balance - KERN ABT 220-4M; id. number: 34/07,
thermometer - enclosed-scale, (0 – 50) °C, 0,1 °C; id. number 34/05.
Used standards are in accordance to standards of Czech Metrology Institute.

We are declaring to our exclusive responsibility, that product, which relates to this declaration responds to the parameters of standard EN ISO 385:2005.

Date of edition: 06. 06. 2017

Confirmed by:

Manager of Quality

170 987 / 2