

Declaration of conformity

Individual

Declaration of conformity given by the supplier of product according to the standard EN ISO IC 17050-1



Establishment TECHNOSKLO s.r.o. Držkov 135 applied and using system of quality according to the DIN EN ISO 9001: 2015, certificate made out by TÜV Rheinland FRG


Sort of product:	1595 / AS ONE - MARK PIPETTE
Nominal volume:	5 ml
Permitted volume error :	± 0,015 ml
Average value of volume:	5,007 ml
Uncertainty of measurement:	± 0,012 ml
Number of charge:	1006

Testing instrument: balance - KERN ABT 220-4M; id. number: 34/07,
thermometer - enclosed-scale, (0 – 50) °C, 0,01 °C; id. number 1/02.
Used standards are in accordance to standards of state metrology .

We are declaring to our exclusive responsibility, that product, which relates to this declaration responds to the parameters of standard: ČSN EN ISO 648:2009.

Date of edition: 26. 10. 2017

Confirmed by:


Manager of Quality

171 725 / 4