

# Checking record



Establishment TECHNOSKLO s.r.o. Držkov 135 applied and using system of quality according to the DIN EN ISO 9 001: 2015, certificate made out by TÜV Rhainland FRG

Sort of product: **632 435 376 725 - BURETTE WITH LATERAL STOPCOCK, GLASS KEY,  
WITH SCHELLBACH STRIPE, QUALICOLOR, CLASS AS**

Nominal volume: **50 ml**


Division: **0,1 ml**

Permitted volume error: **± 0,05 ml**

Testing instrument: balance – KERN ABT 220-4M; id. number: 34 / 07,  
thermometer – enclosed-scale,(0-50)°C, 0,01°C; id. number 1/ 02.  
Used standards are in accordance to standards of State Metrology.

We are declaring to our exclusive responsibility, that product, which relates to this declaration responds to the parameters of standard: ČSN EN ISO 385:2005.

Date of edition: 12. 12. 2017

Confirmed by:   
manager of quality 171 205 / 7